



Translation conformal to original text in Italian. In case of any discrepancies, the Italian version shall prevail



Reg. No	MID 177_F1_03	Revision	00
First issuance	2021-02-12	Last modification	2021-02-12
Page	1 / 1		

## Certificate of conformity based on product verification

We certify that, on the basis of tests and assessment carried out, the following lot of measuring instruments of the Organization :

### **BYP Global – Branding your POS S.L.**

#### **Registered Headquarters**

Avenida San Francisco Javier, numero 9 – Edificio Sevilla 2, planta 4a, modulo 26 – 41018 – Sevilla – Spain

is in compliance with the applicable requirements of D.Lgs. 2 February 2007, No, 22, as modified by D.Lgs. 19 May 2016, n. 84, that transposes Directive 2014/32/UE in Italy, according to Annex II Module F1

#### **Measuring instrument:**

Capacity measure for immediate consumption of beverages

#### **Model:**

Beer Tower 3L

#### **Batch No.:**

OT 20259 (400 Vessels)

#### **Applicable essential requirements:**

Annex I (Essential requirements) / Annex X MI-008 Chapt. II (Specific requirements)

This certificate is subject to compliance with the contractual requirements of Kiwa Cermet Italia and is only valid to the extent identified above. The main features and conditions for approval are specified in the technical Annex that is an integral part of this Certificate of Conformity.

This certificate cannot be reproduced, if not in its integral form

#### **Chief Operating Officer**

*Giampiero Belcredi*

**Kiwa Cermet Italia S.p.A.**  
Società con socio unico,  
soggetta all'attività di  
direzione e coordinamento di  
Kiwa Italia Holding Srl

Via Cadriano, 23  
40057 Granarolo dell'Emilia  
(BO)

Tel +39.051.459.3.111

Fax +39.051.763.382

E-mail: info@kiwacermet.it

**CERMET**



Organismo Notificato n° 0476

Reg. No MID 177\_F1\_03 Revision 00  
First issuance 2021-02-12 Last modification 2021-02-12  
Page 1 / 1

## Technical annex to the certificate of conformity

### Characteristics of the capacity measure:

Vessel with 3 liters of nominal capacity, equipped with a manual system for spilling (Pict. 1)



Pict. 1

### Batch characteristics:

> The batch consists of 400 measures, contained in groups of 4 in:

- 100 carton boxes bearing the batch identification code, numbered from 1 to 100 (red version)

### Technical documentation of reference:

Document CE Marking Beer Tower n. 2 ed. 1 dated 2020-10-27

### Place and date of verification:

Firm Carrer Riu Vinalopó – 23. P.I. Patá del Cid. Apdo. Correos 27 – 46930 Quart de Poblet – Valencia – Spain / 2021-02-09

### Test report:

50 components of the batch were randomly sampled, in compliance with single sampling plan H of level II for tightened inspection, according to ISO 2859-1:2011. The batch resulted acceptable.

Test results are described in test report No. MID 177-20210209-01

### Main reference standards:

OIML R 138:2009

### Security measures:

N.A.



PRD N° 069B



Organismo Notificato n° 0476

